


DECLARATION OF CONFORMITY

Application of Council Directive(s)	2004/108/EC
Standard(s) to which Conformity is Declared	EN61326-1, EN 55022, EN61000-3-2, EN61000-3-3, EN61000-4-2, EN61000-4-3, EN61000-4-4, EN61000-4-5, EN61000-4-6, EN61000-4-11
Manufacturer's Name	Eutech Instruments Pte Ltd
Manufacturer's Address	Blk 55 Ayer Rajah Crescent #04-16/24 Singapore 139949
Type of Equipment	pH/ORP Process Controller

PRODUCT	DESCRIPTION
ECPHCTP0200	Alpha pH 200 pH/ORP Controller
ECPHCTP0190	Alpha pH 190 pH/ORP Controller

ECPHCTP0200 and ECPHCTP0190 are variations of Alpha pH 200, and contain the same electronic components, but differ in software functions. Alpha pH 200 is certified to comply with the above Directives and Standards (Test reference number: G7M20205-0043-E-16).

Manufacturer

Signature  Date May 5, 2008

Full Name V. SRINIDHI

Position DIRECTOR - OPERATIONS





ELECTRONIC TECHNOLOGY SYSTEMS
DR. GENZ GMBH

EC DECLARATION OF CONFORMITY

This certifies that the following designated product

PH/ORP CONTROLLER / TRANSMITTER
MODEL NO. : ALPHA PH200

.....
(Product identification)

complies with the essential protection requirements of Council Directive 89/336/EEC on the approximation of the laws of the Member States relating to electromagnetic compatibility.

This declaration applies to all specimens manufactured in accordance with the attached manufacturing drawings which form part of this declaration.

Assessment of compliance of the product with the requirements relating to electromagnetic compatibility was based on the following standards:

EN 61326-1 : 1997, EN 55022, EN 61000-3-2 /-3
EN 61000-4-2 /-3 /-4 /-5 /-6 /-11

.....
(Identification of regulations / standards)

This declaration is the responsibility of the manufacturer / importer

EUTECH INSTRUMENTS PTE LTD
BLK55, AYER RAJAH CRESCENT, #04-16/24 ,
SINGAPORE 139949, SINGAPORE

.....
(Name / Address)



THIS DOC IS ONLY VALID IN CONNECTION WITH THE TEST REPORT NUMBER : G7M20108-0031-E-16

MANUFACTURER / IMPORTER

TEST LABORATORY

This is the result of test, that was carried out from the submitted type-samples of a product in conformity with the specification of the respective standards.
The certificate holder has the right to fix the CE-mark for EMC on the product complying with the inspection sample.

.....
(Date)

December 11, 2001

.....
(Date)

.....
(Surname, forename)
.....
(Company stamp)

.....
Dr. Genz
.....
Dr. Genz
(Company stamp)



ELECTRONIC TECHNOLOGY SYSTEMS DR. GENZ GMBH

STORKOWER STRASSE 38C, D-15526 REICHENWALDE B. BERLIN, GERMANY, PHONE + 49-33631-88800 FAX + 49-33631-88866