



ELECTRONIC TECHNOLOGY SYSTEMS
DR. GENZ GMBH

EC DECLARATION OF CONFORMITY

This certifies that the following designated product

PH METER
MODEL NO. : PH 5/PH 6

.....
(Product identification)

complies with the essential protection requirements of Council Directive 89/336/EEC on the approximation of the laws of the Member States relating to electromagnetic compatibility.

This declaration applies to all specimens manufactured in accordance with the attached manufacturing drawings which form part of this declaration.

Assessment of compliance of the product with the requirements relating to electromagnetic compatibility was based on the following standards:

EN 50081-1 /1992: EN 55011
EN 50082-1 /1997: EN 61000-4-2/-3

.....
(Identification of regulations / standards)

This declaration is the responsibility of the manufacturer / importer

EUTECH INSTRUMENTS PTE LTD
BLK 55, AYER RAJAH CRESCENT, #04-14/24,
SINGAPORE 139949,

.....
(Name / Address)



THIS DOC IS ONLY VALID IN CONNECTION WITH THE TEST REPORT NUMBER : 12029905

MANUFACTURER / IMPORTER

TEST LABORATORY

This is the result of test, that was carried out from the submitted type-samples of a product in conformity with the specification of the respective standards.
The certificate holder has the right to fix the CE-mark for EMC on the product complying with the inspection sample.

February 12, 1999

.....
(Date)

.....
(Date)

.....
(Surname, forename)

.....
Dr. Genz

.....
(Company stamp)

ELECTRONIC TECHNOLOGY SYSTEMS DR. GENZ GMBH

STORKOWER STRASSE 38C, D-15526 REICHENWALDE B. BERLIN, GERMANY. PHONE + 49-33631-88800 FAX + 49-33631-88866