

Method of Test

The assessment detailed in this document was carried out in accordance with clauses 13.4 (IP6X) and 14.2.7 (IPX7) of IEC 529 : 1989.

Results


Clause No.	Title/Description	Results
13.4	Dust-tight (IP6X)	Complied
14.2.7	Temporary immersion in water test (IPX7)	Complied

Remarks:

- 1) The waterproof handheld meter was submitted for IP67 test.
- 2) The waterproof handheld meter submitted was deemed to comply with the requirement of IP67.



KWOK YEE MENG
ASST. TECHNICAL EXECUTIVE



LAU SENG TAY
PRINCIPAL TECHNICAL EXECUTIVE
ELECTRICAL & ELECTRONICS
TEST CENTRE

/my